

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 29.972 °C                  | 29.8 °C            | -0.172 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 70 °C                      | 70.072 °C                  | 73.6 °C            | 3.528 °C               | ± 0.5 °C                      | ± 1.1E+00 °C                         |
| 30 %HR                     | 28.914 %HR                 | 36.8 %HR           | 7.886 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 51.296 %HR                 | 60.0 %HR           | 8.704 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |