

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.112 °C	30.1 °C	-0.012 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.325 %HR	36.7 %HR	1.375 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.274 %HR	51.7 %HR	1.426 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.4 %HR	-2.048 %HR	± 3.0 %HR	± 1.9E+00 %HR