

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.845 °C	30.3 °C	-0.545 °C	± 1.0 °C	± 6.0E-01 °C
30 %HR	29.946 %HR	35.0 %HR	5.054 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.439 %HR	56.3 %HR	4.861 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.783 %HR	78.1 %HR	3.317 %HR	± 5.0 %HR	± 1.9E+00 %HR