

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.363 °C	30.6 °C	0.237 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	25.4 %HR	-4.659 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.795 %HR	41.8 %HR	-8.995 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.114 %HR	55.9 %HR	-16.214 %HR	± 3.0 %HR	± 1.9E+00 %HR