

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.7 μm	52.5 μm	-0.17 μm	± 5 μm	±	5.8E-01 μm
155.2 μm	155.9 μm	0.71 μm	± 5 μm	±	5.8E-01 μm
257.12 μm	258.22 μm	1.10 μm	± 5 μm	±	5.8E-01 μm
511.72 μm	511.92 μm	0.20 μm	± 5 μm	±	5.8E-01 μm