

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.250 °C	0.7 °C	0.450 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.6 °C	0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.876 °C	34.8 °C	-0.076 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	33.1 %HR	3.041 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.110 %HR	54.4 %HR	-0.710 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.488 %HR	79.8 %HR	-0.688 %HR	± 3.0 %HR	± 1.9E+00 %HR