

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | 0.250 °C                   | 0.2 °C             | -0.050 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.436 °C                  | 20.4 °C            | -0.036 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 35 °C                      | 34.876 °C                  | 34.8 °C            | -0.076 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.059 %HR                 | 32.8 %HR           | 2.741 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 55 %HR                     | 55.110 %HR                 | 57.0 %HR           | 1.890 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 80 %HR                     | 80.488 %HR                 | 79.6 %HR           | -0.888 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |