

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.250 °C	-0.1 °C	-0.350 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.0 °C	-0.436 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.876 °C	34.9 °C	0.024 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.8 %HR	2.741 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.110 %HR	56.0 %HR	0.890 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.488 %HR	79.2 %HR	-1.288 %HR	± 3.0 %HR	± 1.9E+00 %HR