

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.250 °C	-0.2 °C	-0.450 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.613 °C	8.8 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
16 °C	16.565 °C	16.4 °C	-0.165 °C	± 0.5 °C	± 4.0E-01 °C