

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -27 °C	-27.095 °C	-26.7 °C	0.395 °C	± 0.5 °C	± 4.0E-01 °C
-18 °C	-18.271 °C	-18.1 °C	0.171 °C	± 0.5 °C	± 4.0E-01 °C
-13 °C	-13.457 °C	-13.4 °C	0.057 °C	± 0.5 °C	± 4.0E-01 °C