

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.250 °C	0.2 °C	-0.050 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.793 °C	5.2 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.037 °C	10.4 °C	0.363 °C	± 0.5 °C	± 4.0E-01 °C