

Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1	mm	1.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
1.25	mm	1.251	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
1.50	mm	1.501	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
2	mm	2.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
3	mm	3.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
5	mm	5.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
8	mm	8.001	mm	1.000 μm	± 4 μm	±	5.9E-01 μm
10	mm	10.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
15	mm	15.001	mm	1.000 μm	± 4 μm	±	5.8E-01 μm
20	mm	20.001	mm	1.000 μm	± 4 μm	±	5.9E-01 μm
25	mm	25.001	mm	1.000 μm	± 4 μm	±	5.9E-01 μm