

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.845 °C	31.1 °C	0.255 °C	± 1.0 °C	± 6.0E-01 °C
30 %HR	29.599 %HR	26 %HR	-3.599 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	51.439 %HR	49 %HR	-2.439 %HR	± 5 %HR	± 1.9E+00 %HR
75 %HR	74.783 %HR	76 %HR	1.217 %HR	± 5 %HR	± 1.9E+00 %HR