

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.559 °C                  | 10.4 °C            | -0.159 °C              | ± 0.5 °C                      | ± 6.0E-01 °C                         |
| 20 °C                      | 20.245 °C                  | 20.1 °C            | -0.145 °C              | ± 0.5 °C                      | ± 6.0E-01 °C                         |
| 30 °C                      | 29.892 °C                  | 29.9 °C            | 0.008 °C               | ± 0.5 °C                      | ± 6.0E-01 °C                         |
| 40 °C                      | 39.719 °C                  | 39.7 °C            | -0.019 °C              | ± 0.5 °C                      | ± 6.0E-01 °C                         |