

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
10 °C	10.559 °C	10.4 °C	-0.159 °C	± 0.5 °C	± 6.0E-01 °C
20 °C	20.245 °C	20.2 °C	-0.045 °C	± 0.5 °C	± 6.0E-01 °C
30 °C	29.892 °C	29.9 °C	0.008 °C	± 0.5 °C	± 6.0E-01 °C
40 °C	39.719 °C	39.8 °C	0.081 °C	± 0.5 °C	± 6.0E-01 °C