

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.559 °C	10.7 °C	0.141 °C	± 0.5 °C	± 6.0E-01 °C
20 °C	20.245 °C	20.3 °C	0.055 °C	± 0.5 °C	± 6.0E-01 °C
30 °C	29.892 °C	29.8 °C	-0.092 °C	± 0.5 °C	± 6.0E-01 °C
40 °C	39.719 °C	39.7 °C	-0.019 °C	± 0.5 °C	± 6.0E-01 °C