

| Nominal (<i>Nominal</i>) | Patrón (<i>Standard</i>) | UBP (<i>UUT</i>) | Error (<i>Error</i>) | Exactitud (<i>Accuracy</i>) | Incertidumbre (<i>Uncertainty</i>) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---------------------------------------------|
| 20 °C | 19.945 °C | 20 °C | 0.055 °C | ± 0.5 °C | ± 6.0E-01 °C |
| 30 °C | 39.849 °C | 40 °C | 0.151 °C | ± 0.5 °C | ± 6.0E-01 °C |
| 40 %HR | 40.171 %HR | 41 %HR | 0.829 %HR | ± 3 %HR | ± 1.9E+00 %HR |
| 70 %HR | 69.670 %HR | 70 %HR | 0.330 %HR | ± 3 %HR | ± 1.9E+00 %HR |