

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	19.945 °C	20 °C	0.055 °C	± 0.5 °C	± 6.0E-01 °C
30 °C	29.822 °C	30 °C	0.178 °C	± 0.5 °C	± 6.0E-01 °C
40 %HR	40.171 %HR	37 %HR	-3.171 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.670 %HR	68 %HR	-1.670 %HR	± 3 %HR	± 1.9E+00 %HR