

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	19.945 °C	20 °C	0.055 °C	± 0.5 °C	± 6.0E-01 °C
30 °C	29.822 °C	30 °C	0.178 °C	± 0.5 °C	± 6.0E-01 °C
40 %HR	40.171 %HR	41 %HR	0.829 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.670 %HR	68 %HR	-1.670 %HR	± 3 %HR	± 1.9E+00 %HR