

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.30 μm	23.0 μm	-0.30 μm	± 2.5 μm	±	5.8E-01 μm
51.60 μm	51.8 μm	0.20 μm	± 2.5 μm	±	5.8E-01 μm
124.60 μm	125.2 μm	0.60 μm	± 2.5 μm	±	5.8E-01 μm
250.30 μm	251.1 μm	0.80 μm	± 2.5 μm	±	5.8E-01 μm
486.00 μm	484.7 μm	-1.30 μm	± 2.5 μm	±	5.8E-01 μm
963.00 μm	960.5 μm	-2.50 μm	± 2.5 μm	±	6.0E-01 μm