

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 25 °C                    | 24.848 °C                | 24.9 °C          | 0.052 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| # 25 °C                  | 24.848 °C                | 25.0 °C          | 0.152 °C             | ± 0.5 °C                    | ± 7.4E-02 °C                       |
| # 50 °C                  | 49.917 °C                | 50.4 °C          | 0.483 °C             | ± 0.5 °C                    | ± 7.4E-02 °C                       |
| # 70 °C                  | 69.972 °C                | 70.8 °C          | 0.828 °C             | ± 0.5 °C                    | ± 7.4E-02 °C                       |
| 25 %HR                   | 23.413 %HR               | 26.7 %HR         | 3.287 %HR            | ± 3 %HR                     | ± 1.9E+00 %HR                      |
| 50 %HR                   | 48.975 %HR               | 52.5 %HR         | 3.525 %HR            | ± 3 %HR                     | ± 1.9E+00 %HR                      |
| 70 %HR                   | 69.425 %HR               | 73.0 %HR         | 3.575 %HR            | ± 3 %HR                     | ± 1.9E+00 %HR                      |