

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49966	0.340	± 0.8	±	4.3E-01 mm						
0 a 1	0.99965	0.350	± 1	±	4.3E-01 mm						
0 a 1.5	1.49959	0.410	± 1.2	±	4.8E-01 mm						
0 a 2	1.99975	0.250	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49937	0.630	± 1.6	±	4.8E-01 mm						
0 a 3	2.99943	0.570	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49940	0.600	± 2	±	4.8E-01 mm						
0 a 4	3.99903	0.970	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49908	0.920	± 2.4	±	4.8E-01 mm						
0 a 5	4.99909	0.910	± 2.6	±	4.8E-01 mm						