

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.64 μm	-0.01 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	155.2 μm	-0.03 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	257.2 μm	0.08 μm	± 2.5 μm	±	5.8E-01 μm
511.72 μm	512.2 μm	0.48 μm	± 2.5 μm	±	5.8E-01 μm