

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.09997	0.030	± 0.2	±	4.3E-01 mm						
0 a 0.2	0.19993	0.070	± 0.2	±	4.3E-01 mm						
0 a 0.3	0.29993	0.070	± 0.2	±	4.3E-01 mm						
0 a 0.4	0.39995	0.050	± 0.2	±	4.3E-01 mm						
0 a 0.5	0.49995	0.050	± 0.3	±	4.3E-01 mm						
0 a 0.6	0.59997	0.030	± 0.3	±	4.3E-01 mm						
0 a 0.7	0.69999	0.010	± 0.3	±	4.3E-01 mm						
0 a 0.8	0.80001	-0.010	± 0.3	±	4.3E-01 mm						
0 a 0.9	0.90005	-0.050	± 0.3	±	4.3E-01 mm						
0 a 1	1.00010	-0.100	± 0.4	±	4.3E-01 mm						