

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.7 μm	52.8 μm	0.15 μm	± 5 μm	±	5.8E-01 μm
155.2 μm	155.2 μm	0.01 μm	± 5 μm	±	5.8E-01 μm
257.12 μm	257.24 μm	0.12 μm	± 5 μm	±	5.8E-01 μm
511.72 μm	512.20 μm	0.48 μm	± 5 μm	±	5.8E-01 μm