

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
52.65 μm	52.62 μm	-0.03 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	154.8 μm	-0.43 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	257.0 μm	-0.12 μm	± 2.5 μm	±	5.8E-01 μm
511.72 μm	512.0 μm	0.28 μm	± 2.5 μm	±	5.8E-01 μm