

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 1.005 mm                 | 1.023 mm         | 18 μm                | ± 50.25 μm               | ±                                  | 5.8E-01 μm |
| 1.5 mm                   | 1.518 mm         | 18 μm                | ± 75. μm                 | ±                                  | 5.8E-01 μm |
| 2 mm                     | 2.019 mm         | 19 μm                | ± 100 μm                 | ±                                  | 5.8E-01 μm |
| 3 mm                     | 4.028 mm         | 1028 μm              | ± 150. μm                | ±                                  | 5.8E-01 μm |
| 4 mm                     | 5.031 mm         | 1031 μm              | ± 200 μm                 | ±                                  | 5.8E-01 μm |