

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 13 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 13 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	0 µm	± 13 µm	±	5.8E+00 µm
2.00 mm	2.00 mm	0 µm	± 13 µm	±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm	± 13 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 13 µm	±	5.8E+00 µm
6.00 mm	5.93 mm	-70 µm	± 13 µm	±	5.8E+00 µm
8.00 mm	7.92 mm	-80 µm	± 13 µm	±	5.8E+00 µm
10.00 mm	9.87 mm	-130 µm	± 13 µm	±	5.8E+00 µm