

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.01	0.00989	0.110	± 0.6	±	4.3E-01 mm						
0 a 0.02	0.01984	0.160	± 0.6	±	4.3E-01 mm						
0 a 0.03	0.02978	0.220	± 0.6	±	4.3E-01 mm						
0 a 0.04	0.03989	0.110	± 0.6	±	4.3E-01 mm						
0 a 0.05	0.04981	0.190	± 0.6	±	4.3E-01 mm						
0 a 0.06	0.05986	0.140	± 0.6	±	4.3E-01 mm						
0 a 0.07	0.06987	0.130	± 0.6	±	4.3E-01 mm						
0 a 0.08	0.07978	0.220	± 0.6	±	4.3E-01 mm						
0 a 0.09	0.08981	0.190	± 0.6	±	4.3E-01 mm						
0 a 0.1	0.09978	0.220	± 0.6	±	4.3E-01 mm						
0 a 0.11	0.10979	0.210	± 0.6	±	4.3E-01 mm						