

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.01	0.01006	-0.060	± 0.6	±	4.3E-01 mm						
0 a 0.02	0.02019	-0.190	± 0.6	±	4.3E-01 mm						
0 a 0.03	0.03025	-0.250	± 0.6	±	4.3E-01 mm						
0 a 0.04	0.04019	-0.190	± 0.6	±	4.3E-01 mm						