

Clase de Exactitud (Accuracy Class): Clase II (Class II)

---

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09967	0.330	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19986	0.140	± 0.6	± 4.3E-01 mm					
0 a 0.4	0.39968	0.320	± 0.7	± 4.3E-01 mm					
0 a 0.6	0.59951	0.490	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.79966	0.340	± 0.9	± 4.3E-01 mm					
0 a 1	0.99980	0.200	± 1	± 4.3E-01 mm					
0 a 1.2	1.19965	0.350	± 1	± 4.8E-01 mm					
0 a 1.4	1.40001	-0.010	± 1.1	± 4.8E-01 mm					
0 a 1.6	1.60006	-0.060	± 1.2	± 4.8E-01 mm					
0 a 1.8	1.80017	-0.170	± 1.3	± 4.8E-01 mm					
0 a 2	2.00017	-0.170	± 1.4	± 4.8E-01 mm					
0 a 2.34	2.33945	0.550	± 1.5	± 4.8E-01 mm					

---