

---

<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
59.96 RPM	60 RPM	0.04 RPM	± 60.1 RPM	± 1.3E-01 RPM

---