

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	27.4 %HR	-3.170 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	53.6 %HR	2.917 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	73.7 %HR	3.252 %HR	± 3.0 %HR	± 1.9E+00 %HR