

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
28 °C	28.057 °C	28.2 °C	0.143 °C	± 0.6 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.3 °C	0.137 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	27.4 %HR	-3.170 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	53.6 %HR	2.917 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	73.70 %HR	3.252 %HR	± 3.0 %HR	± 1.9E+00 %HR