

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.313 °C	30.0 °C	-0.313 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	36.0 %HR	0.317 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	52.0 %HR	1.982 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.0 %HR	-1.448 %HR	± 5.0 %HR	± 1.9E+00 %HR