

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99938	0.620	± 1	±	#REF! mm						
0 a 2	1.99979	0.210	± 1.4	±	#REF! mm						
0 a 3	2.99953	0.470	± 1.8	±	#REF! mm						
0 a 4	3.99990	0.100	± 2.2	±	#REF! mm						
0 a 5	4.99980	0.200	± 2.6	±	#REF! mm						
0 a 6	6.00007	-0.070	± 3	±	#REF! mm						
0 a 7	6.99991	0.090	± 3.4	±	#REF! mm						
0 a 8	8.00018	-0.180	± 3.8	±	#REF! mm						
0 a 9	8.99990	0.100	± 4.2	±	#REF! mm						
0 a 10	10.00017	-0.170	± 4.6	±	#REF! mm						
0 a 11	10.99982	0.180	± 5	±	#REF! mm						
0 a 12	12.00025	-0.250	± 5.4	±	#REF! mm						
0 a 13	13.00020	-0.200	± 5.8	±	#REF! mm						
0 a 14	14.00051	-0.510	± 6.2	±	#REF! mm						
0 a 15	15.00001	-0.010	± 6.6	±	#REF! mm						
0 a 16	16.00033	-0.330	± 7	±	#REF! mm						
0 a 17	17.00004	-0.040	± 7.4	±	#REF! mm						
0 a 18	18.00013	-0.130	± 7.8	±	#REF! mm						
0 a 19	19.00034	-0.340	± 8.2	±	#REF! mm						
0 a 20	20.00066	-0.660	± 8.6	±	#REF! mm						
0 a 21	21.00054	-0.540	± 9	±	#REF! mm						
0 a 22	22.00092	-0.920	± 9.4	±	#REF! mm						
0 a 23	23.00041	-0.410	± 9.8	±	#REF! mm						
0 a 24	24.00078	-0.780	± 10.2	±	#REF! mm						
0 a 25	25.00052	-0.520	± 10.6	±	#REF! mm						
0 a 26	26.00090	-0.900	± 11	±	#REF! mm						
0 a 27	27.00061	-0.610	± 11.4	±	#REF! mm						
0 a 28	28.00110	-1.100	± 11.8	±	#REF! mm						
0 a 29	29.00093	-0.930	± 12.2	±	#REF! mm						
0 a 30	30.00146	-1.460	± 12.6	±	#REF! mm						