

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99938	0.620	± 1	± 4.3E-01 mm					
0 a 2	1.99979	0.210	± 1.4	± 4.8E-01 mm					
0 a 3	2.99953	0.470	± 1.8	± 4.8E-01 mm					
0 a 4	3.99990	0.100	± 2.2	± 4.8E-01 mm					
0 a 5	4.99980	0.200	± 2.6	± 4.8E-01 mm					
0 a 6	6.00007	-0.070	± 3	± 3.6E-01 mm					
0 a 7	6.99991	0.090	± 3.4	± 3.6E-01 mm					
0 a 8	8.00018	-0.180	± 3.8	± 3.6E-01 mm					
0 a 9	8.99990	0.100	± 4.2	± 3.6E-01 mm					
0 a 10	10.00017	-0.170	± 4.6	± 3.6E-01 mm					
0 a 11	10.99982	0.180	± 5	± 3.6E-01 mm					
0 a 12	12.00025	-0.250	± 5.4	± 3.6E-01 mm					
0 a 13	13.00020	-0.200	± 5.8	± 3.6E-01 mm					
0 a 14	14.00051	-0.510	± 6.2	± 3.6E-01 mm					
0 a 15	15.00001	-0.010	± 6.6	± 3.6E-01 mm					
0 a 16	16.00033	-0.330	± 7	± 3.6E-01 mm					
0 a 17	17.00004	-0.040	± 7.4	± 3.6E-01 mm					
0 a 18	18.00013	-0.130	± 7.8	± 3.6E-01 mm					
0 a 19	19.00034	-0.340	± 8.2	± 3.6E-01 mm					
0 a 20	20.00066	-0.660	± 8.6	± 3.6E-01 mm					
0 a 21	21.00054	-0.540	± 9	± 5.5E-01 mm					
0 a 22	22.00092	-0.920	± 9.4	± 5.5E-01 mm					
0 a 23	23.00041	-0.410	± 9.8	± 5.5E-01 mm					
0 a 24	24.00078	-0.780	± 10.2	± 5.5E-01 mm					
0 a 25	25.00052	-0.520	± 10.6	± 5.5E-01 mm					
0 a 26	26.00090	-0.900	± 11	± 5.5E-01 mm					
0 a 27	27.00061	-0.610	± 11.4	± 5.5E-01 mm					
0 a 28	28.00110	-1.100	± 11.8	± 5.5E-01 mm					
0 a 29	29.00093	-0.930	± 12.2	± 5.5E-01 mm					
0 a 30	30.00146	-1.460	± 12.6	± 5.5E-01 mm					