

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49968	0.320	± 0.8	± 4.3E-01 mm					
0 a 1	0.99980	0.200	± 1	± 4.3E-01 mm					
0 a 1.5	1.49994	0.060	± 1.2	± 4.8E-01 mm					
0 a 2	1.99950	0.500	± 1.4	± 4.8E-01 mm					
0 a 2.5	2.49951	0.490	± 1.6	± 4.8E-01 mm					
0 a 3	2.99929	0.710	± 1.8	± 4.8E-01 mm					
0 a 3.5	3.49951	0.490	± 2	± 4.8E-01 mm					
0 a 4	3.99923	0.770	± 2.2	± 4.8E-01 mm					
0 a 4.5	4.49932	0.680	± 2.4	± 4.8E-01 mm					
0 a 5	4.99948	0.520	± 2.6	± 4.8E-01 mm					
