

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.263 °C	30.8 °C	0.537 °C	± 0.8 °C	± 4.0E-01 °C
30 %HR	30.672 %HR	29.8 %HR	-0.872 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	50.1 %HR	-0.583 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.8 %HR	-2.648 %HR	± 4.0 %HR	± 1.9E+00 %HR