

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.653 °C	16.1 °C	0.447 °C	± 0.8 °C	± 4.0E-01 °C
20 °C	20.155 °C	20.5 °C	0.345 °C	± 0.8 °C	± 4.0E-01 °C
30 %HR	30.672 %HR	29.8 %HR	-0.872 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	50.1 %HR	-0.583 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.80 %HR	-2.648 %HR	± 4.0 %HR	± 1.9E+00 %HR