

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 15 °C             | 15.392 °C         | 15.1 °C   | -0.292 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.550 °C         | 25.1 °C   | -0.450 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 32 °C             | 31.867 °C         | 31.5 °C   | -0.367 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 20 %HR            | 20.652 %HR        | 23.0 %HR  | 2.348 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.304 %HR        | 53.2 %HR  | 2.896 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 73.2 %HR  | 2.752 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |