

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.392 °C	15.5 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	24.9 °C	-0.650 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.867 °C	31.9 °C	0.033 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.652 %HR	22.2 %HR	1.548 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.304 %HR	52.0 %HR	1.696 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.8 %HR	1.352 %HR	± 3.0 %HR	± 1.9E+00 %HR