

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.392 °C	15.3 °C	-0.092 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.3 °C	-0.250 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.867 °C	32.1 °C	0.233 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.652 %HR	23.2 %HR	2.548 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.304 %HR	49.2 %HR	-1.104 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.6 %HR	-1.848 %HR	± 3.0 %HR	± 1.9E+00 %HR