

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.1 °C	0.058 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.098 %HR	36.0 %HR	1.902 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.366 %HR	52.3 %HR	1.934 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.834 %HR	71.7 %HR	1.866 %HR	± 3.0 %HR	± 1.9E+00 %HR