

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.0 °C	-0.042 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.098 %HR	35.1 %HR	1.002 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.816 %HR	51.7 %HR	0.884 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.834 %HR	67.4 %HR	-2.434 %HR	± 3.0 %HR	± 1.9E+00 %HR