

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.900 °C	3.9 °C	0.000 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	7.8 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	30.0 °C	-0.042 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.621 %HR	52.0 %HR	1.379 %HR	± 5.0 %HR	± 1.9E+00 %HR