

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.900 °C	4.0 °C	0.100 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	8.2 °C	0.178 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	29.8 °C	-0.242 °C	± 0.5 °C	± 4.0E-01 °C