

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.900 °C	3.1 °C	-0.800 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	6.5 °C	-1.522 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	27.9 °C	-2.142 °C	± 0.5 °C	± 4.0E-01 °C