

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
4 °C	3.900 °C	3.2 °C	-0.700 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	7.0 °C	-1.022 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	27.9 °C	-2.142 °C	± 0.5 °C	± 4.0E-01 °C