

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	3.900 °C	4.0 °C	0.100 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	8.2 °C	0.178 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	29.9 °C	-0.142 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.621 %HR	53.6 %HR	2.979 %HR	± 5.0 %HR	± 1.9E+00 %HR